

Mixed Analyte Performance Evaluation Program

Statistical Summary

Sample ID: MAPEP-05-GrF13

Analyte	T(1)	A(2)	Grand Mean	Std. Dev.	Reference Value	Analyte Text	Acceptance Limits	Units
Gross alpha	57	54	0.11	0.03	0.232		>0.0 - 0.46	Bq/sample
Gross beta	58	53	0.34	0.05	0.297		0.15 - 0.45	Bq/sample

Gross Alpha Flags:

A = Result acceptable, Bias \leq +/- 100% with a statistically positive result at two standard deviations (Result/Uncertainty > 2 , i.e., the range encompassing the result, plus or minus the total uncertainty at two standard deviations, does not include zero).

N = Result not acceptable, Bias $>$ +/- 100% or the reported result is not statistically positive at two standard deviations (Result/Uncertainty ≤ 2 , i.e., the range encompassing the result, plus or minus the total uncertainty at two standard deviations, includes zero).

Gross Beta Flags:

A = Result acceptable, Bias \leq +/- 50% with a statistically positive result at two standard deviations (Result/Uncertainty > 2 , i.e., the range encompassing the result, plus or minus the total uncertainty at two standard deviations, does not include zero).

N = Result not acceptable, Bias $>$ +/- 50% or the reported result is not statistically positive at two standard deviations (Result/Uncertainty ≤ 2 , i.e., the range encompassing the result, plus or minus the total uncertainty at two standard deviations, includes zero).

Note: Outliers are excluded from the statistical summary.
Outliers are defined above.
(1) T = Total Number of Laboratories Reporting Analyte.
(2) A = Number of Laboratories with 'Acceptable' Performance.