

Sample Statistical Summary

MAPEP-09-GrW20

Analyte	T(1)	A(2)	Grand(3) Mean	Std Dev	Ref Value	Ref Unc	Acceptance Range
Radiological							Units: (Bq/L)
Gross alpha	58	53	0.655	0.198	0.635	0.011	>0.0 - 1.270
Gross beta	58	52	1.38	0.21	1.27	0.02	0.64 - 1.91

Note: (1) T = Total number of laboratories reporting analyte.
(2) A = Number of laboratories with 'Acceptable' performance.

Gross Alpha Criteria:

Acceptable Result: Bias \leq +/- 100% with a statistically positive result at two standard deviations (Result/Uncertainty > 2 , i.e., the range encompassing the result, plus or minus the total uncertainty at two standard deviations, does not include zero).

Gross Beta Criteria:

Acceptable Result: Bias \leq +/- 50% with a statistically positive result at two standard deviations (Result/Uncertainty > 2 , i.e., the range encompassing the result, plus or minus the total uncertainty at two standard deviations, does not include zero).