

Sample Statistical Summary

MAPEP-09-GrW21

Analyte	T(1)	A(2)	Grand(3) Mean	Std Dev	Ref Value	Ref Unc	Acceptance Range
Radiological							Units: (Bq/L)
Gross alpha	60	59	1.052	0.369	1.047	0.011	>0.0 - 2.094
Gross beta	58	55	7.47	1.22	7.53	0.01	3.77 - 11.30

Note: (1) T = Total number of laboratories reporting analyte.
(2) A = Number of laboratories with 'Acceptable' performance.

Gross Alpha Criteria:

Acceptable Result: Bias \leq +/- 100% with a statistically positive result at two standard deviations (Result/Uncertainty > 2 , i.e., the range encompassing the result, plus or minus the total uncertainty at two standard deviations, does not include zero).

Gross Beta Criteria:

Acceptable Result: Bias \leq +/- 50% with a statistically positive result at two standard deviations (Result/Uncertainty > 2 , i.e., the range encompassing the result, plus or minus the total uncertainty at two standard deviations, does not include zero).